



Volume Range	Colour Code	Channels	Test Volume (µL)	Inaccuracy (%)	Imprecision (%)
--------------	-------------	----------	------------------	----------------	-----------------

Thermo - Finn timer Pipette Focus Fix Volume Pipette

1	1 µL	Pink	1	± 4	≤ 4
2	2 µL	Pink	2	± 2.5	≤ 2
3	5 µL	Pink	5	± 1.4	≤ 1.4
4	10 µL	Yellow	10	± 0.9	≤ 0.8
5	20 / 25 µL	Yellow	20 / 25	± 0.6	≤ 0.5
6	50 µL	Yellow	50	± 0.6	≤ 0.4
7	100 / 200 / 250 µL	Yellow	100 / 200 / 250	± 0.4	≤ 0.3
8	500 / 1000 µL	Blue	500 / 1000	± 0.3	≤ 0.3
9	2000 / 3000 µL	Green	2000 / 3000	± 0.3	≤ 0.2
10	5000 / 10000 µL	Red	5000 / 10000	± 0.3	≤ 0.2

Thermo - Finn timer Pipette Focus Single Channel Pipette

1	0.3-3 µL	Pink	1	± 10_2.5	≤ 8.2
2	0.5-5 µL	Pink	1	± 6_1	≤ 5.0.8
3	1-10 µL	Pink	1	± 2.5_1	≤ 2_0.5
4	1-10 µL	Yellow	1	± 3.5_1	≤ 3_0.5
5	3-30 µL	Yellow	1	± 3.5_0.8	≤ 3_0.4
6	5-50 µL	Yellow	1	± 3_0.6	≤ 2.5_0.3
7	10-100 µL	Yellow	1	± 3_0.8	≤ 1_0.2
8	30-300 µL	Orange	1	± 1.5_0.6	≤ 0.6_0.2
9	100-1000 µL	Blue	1	± 1_0.6	≤ 0.6_0.2
10	500-5000 µL	Green	1	± 2_0.5	≤ 0.8_0.2
11	1-10 mL	Red	1	± 2_0.5	≤ 0.8_0.2

Thermo - Finn timer Pipette Focus Multi Channel Pipette

1	1-10 µL	Pink	8_12	± 12_2.4	≤ 8_1.6
2	5-50 µL	Yellow	8_12	± 5_1.5	≤ 2_0.7
3	30-300 µL	Orange	8_12	± 5_1	≤ 2_0.3
4	5-50 µL	Turquoise	16	± 5_1.5	≤ 2_0.7





Volume Range	Colour Code	Channels	Test Volume (μL)	Inaccuracy (%)	Imprecision (%)
--------------	-------------	----------	------------------	----------------	-----------------

Thermo - Finnpiquette Digital & Colour Single Channel Pipette

1	0.2-2 μL	Pink	1	0.2_2	± 12_2.5	≤ 10_2
2	0.5-10 μL	Pink	1	1_10	± 3.5_1	≤ 3_0.5
3	0.5-10 μL	Yellow	1/1	1_10	± 3.5_1	≤ 3_0.8
4	2-20 μL	Yellow	1	2_20	± 3_1	≤ 2.5_0.4
5	5-50 μL	Yellow	1/1	5_50	± 3_0.6	≤ 2.5_0.3
6	10-100 μL	Yellow	1	10_100	± 3_0.8	≤ 1_0.2
7	20-200 μL	Yellow	1/1	20_200	± 1.8_0.6 / 3_0.6	≤ 0.7_0.2 / 1.5_0.3
8	200-1000 μL	Blue	1	200_1000	± 0.9_0.6	≤ 0.3_0.2
9	100-1000 μL	Blue	1/1	100_1000	± 1_0.6 / 1.5_0.5	≤ 0.6_0.2
10	500-5000 μL	Green	1/1	1000_5000	± 0.8_0.5 / 1.5_0.5	≤ 0.3_0.2 / 0.5_0.2
11	2-10 mL	Red	1	2000_10000	± 1_0.5	≤ 0.3_0.2

Thermo - Finnpiquette Digital & Colour Multi Channel Pipette

1	1-10 μL	Pink	8_12	1_10	± 12_2.4	≤ 8_1.6
2	5-50 μL	Yellow	8_12 / 4_8_12	5_50	± 5_1.5	≤ 2_0.7
3	50-300 μL	Orange	8_12 / 4_8_12	50_300	± 4.6_1	≤ 1.5_0.3
4	5-50 μL	Turquoise	16 / 16	5_50	± 5_1.5	≤ 2_0.7

Thermo - Matrix Equalizer

1	0.5-12.5 μL	Red	8_12	5_40	± 2.5	≤ 2
2	1-30 μL	Purple	8_12	40_200	± 2	≤ 1.7
3	2-125 μL	Yellow	8_12	200-1000	± 2	≤ 1
4	5-250 μL	Blue	8	1000-5000	± 2.5	≤ 0.7
5	15-1250 μL	Green	8	2000-10000	± 1.5	≤ 0.6





Volume Range	Colour Code	Channels	Test Volume (µL)	Inaccuracy (%)	Imprecision (%)
--------------	-------------	----------	------------------	----------------	-----------------

Thermo - Finn timer Single Channel Pipette

1	1-10 µL	Pink	1	1_10	± 3.5_1	≤ 3_0.5
2	1-10 µL	Yellow	1	1_10	± 7_1	≤ 6_0.5
3	5-50 µL	Turquoise	1	5_50	± 3_0.8	≤ 2.5_0.3
4	5-50 µL	Yellow	1	5_50	± 3_0.8	≤ 2.5_0.3
5	10-100 µL	Yellow	1	10_100	± 3_0.8	≤ 1_0.2
6	30-300 µL	Orange	1	30_300	± 3_0.6	≤ 0.7_0.2
7	100-1000 µL	Blue	1	100_1000	± 3_0.6	≤ 0.6_0.2
8	500-5000 µL	Green	1	500_5000	± 3_0.6	≤ 0.8_0.2
9	1-10 mL	Red	1	1000_10000	± 3_0.6	≤ 0.8_0.2

Thermo - Finn timer Multi Channel Pipette

1	1-10 µL	Pink	8_12	1_10	± 12_2.4	≤ 8_1.6
2	5-50 µL	Yellow	8_12	5_50	± 5_1.5	≤ 2_0.7
3	30-300 µL	Orange	8_12	30_300	± 3_1	≤ 1_0.3
4	5-50 µL	Turquoise	16	5_50	± 5_1.5	≤ 2_0.7
5	100-1200 µL	Turquoise	8	100_1200	± 3_1	≤ 0.9_0.2

Thermo - Matrix Hybrid Single Channel Pipette

1	0.5-12.5 µL	Pink	1	0.5_12.5	± 3.33_1	≤ 1.8_0.6
2	2-30 µL	Yellow	1	2_30	± 2.8_1	< 1.6_0.5
3	10-125 µL	Yellow	1	10_125	± 2.67_1	≤ 1.12_0.4
4	20-300 µL	Orange	1	20_300	± 2.4_0.9	≤ 1_0.3
5	100-1250 µL	Turquoise	1	100_1250	± 2_0.8	≤ 0.64_0.2

Thermo - Matrix Hybrid Multi Channel Pipette

1	0.5-12.5 µL	Pink	8_12	0.5_12.5	± 6.7_2	≤ 5_1.2
2	2-30 µL	Yellow	8_12	2_30	± 5.6_2	< 3.2_1
3	20-300 µL	Orange	8_12	20_300	± 4.8_1	≤ 2_0.3





	Volume Range	Colour Code	Channels	Test Volume (µL)	Inaccuracy (%)	Imprecision (%)
Thermo - Finnepipette F1 & F2 Fix Volume Pipette						
1	1 µL	Pink	1/1	1	± 4	≤ 4
2	5 µL	Pink	1/1	5	± 1.4	≤ 1.4
3	10 µL	Yellow	1/1	10	± 0.9	≤ 0.8
4	25 µL	Yellow	1/1	25	± 0.6	≤ 0.5
5	50 µL	Yellow	1/1	50	± 0.6	≤ 0.4
6	100 / 250 µL	Yellow	1/1	100 / 250	± 0.4	≤ 0.3
7	500 / 1000 µL	Blue	1/1	500 / 1000	± 0.3	≤ 0.3
8	2000 µL	Blue	1/1	2000	± 0.3	≤ 0.2
9	3000 / 5000 µL	Green	1/1	3000 / 5000	± 0.3	≤ 0.2
10	10 mL	Red	1/1	10000	± 0.3	≤ 0.2

Thermo - Finnepipette F1 & F2 Single Channel Pipette						
1	0.2-2 µL	Pink	1/1	0.2_1_2	± 12_4_2.5	≤ 10_3.5_2
2	0.5-5 µL	Pink	1/1	0.5_2.5_5	± 6_2.5_1 / 6_2.5_1.5	≤ 5_1.5_1
3	1-10 µL	Pink	1/1	1_5_10	± 2.5_1.5_1	≤ 2_0.8_0.5
4	1-10 µL	Yellow	1/1	1_5_10	± 3.5_1.5_1	≤ 3_0.8_0.8
5	2-20 µL	Yellow	1/1	2_10_20	± 3_1.5_1	≤ 2.5_0.6_0.4
6	2-20 µL	Turquoise	1/1	2_10_20	± 3_1.5_1	≤ 2.5_0.6_0.4
7	10-100 µL	Yellow	1/1	10_50_100	± 3_1.2_0.8	≤ 1_0.4_0.2
8	20-200 µL	Yellow	1/1	20_100_200	± 1.8_1_0.6	≤ 0.7_0.4_0.2
9	30-300 µL	Orange	1	30_150_300	± 1.5_1_0.6	≤ 0.6_0.4_0.2
10	100-1000 µL	Blue	1/1	100_500_1000	± 1_0.8_0.6	≤ 0.6_0.3_0.2
11	500-5000 µL	Green	1/1	500_2500_5000	± 2_0.7_0.5	≤ 0.8_0.3_0.2
12	1-10 mL	Red	1/1	1000_5000_10000	± 2_0.8_0.5	≤ 0.8_0.3_0.2





Volume Range	Colour Code	Channels	Test Volume (μL)	Inaccuracy (%)	Imprecision (%)
--------------	-------------	----------	------------------	----------------	-----------------

Thermo - Finnepipette F3 Fix Volume Pipette

1	1 μL	Pink	1	± 4	≤ 4
2	5 μL	Pink	5	± 1.4	≤ 1.4
3	10 μL	Yellow	10	± 0.9	≤ 0.8
4	25 μL	Yellow	25	± 0.6	≤ 0.5
5	50 μL	Yellow	50	± 0.6	≤ 0.4
6	100 μL	Yellow	100	± 0.4	≤ 0.3
7	250 μL	Blue	250	± 0.4	≤ 0.3
8	500 μL	Blue	500	± 0.3	≤ 0.3
9	1000 μL	Blue	1000	± 0.3	≤ 0.3

Thermo - Finnepipette F3 Single Channel Pipette

1	1-10 μL	Pink	1	1_5_10	± 2.5_2_1	≤ 2_1_0.5
2	1-10 μL	Yellow	1	1_5_10	± 3.5_2_1	≤ 3_1.3_0.8
3	2-20 μL	Yellow	1	2_10_20	± 3_1.5_1	≤ 2.5_0.7_0.4
4	2-20 μL	Turquoise	1	2_10_20	± 3_1.5_1	≤ 2.5_0.7_0.4
5	10-100 μL	Yellow	1	10_50_100	± 3_1.2_0.8	≤ 1_0.4_0.2
6	20-200 μL	Yellow	1	20_100_200	± 3_1_0.6	≤ 1.5_0.5_0.3
7	100-1000 μL	Blue	1	100_500_1000	± 1.5_1_0.5	≤ 0.6_0.4_0.2
8	1-10 mL	Red	1	1000_5000_10000	± 2_0.8_0.5	≤ 0.8_0.3_0.2

Thermo - Finnepipette F1 & F2 Multi Channel Pipette

1	1-10 μL	Pink	8_12	1_5_10	± 12_4_2.4	≤ 8_3_1.6
2	1-10 μL	Purple	16	1_5_10	± 12_4_2.4	≤ 8_3_1.6
3	5-50 μL	Yellow	8_12	5_25_50	± 5_2.5_1.5	≤ 2_1.2_0.7
4	10-100 μL	Yellow	8_12	10_50_100	± 5_2.5_1.3	≤ 2_1.2_0.5
5	30-300 μL	Orange	8_12	30_150_300	± 5_1.5_1	≤ 2_0.5_0.3
6	5-50 μL	Turquoise	16	5_25_50	± 5_2.5_1.5	≤ 2_1.2_0.7





Volume Range	Colour Code	Channels	Test Volume (µL)	Inaccuracy (%)	Imprecision (%)
--------------	-------------	----------	------------------	----------------	-----------------

Thermo - Finn timer

1	10-50 µL	Finntip	500 µL	10_50	± 1_1	≤ 1_0.6
2	25-125 µL	Finntip	1250 µL	25_125	± 0.8_0.8	≤ 0.8_0.5
3	50-250 µL	Finntip	2500 µL	50_250	± 0.8_0.8	≤ 0.7_0.3
4	100-500 µL	Finntip	5000 µL	100_500	± 0.8_0.8	≤ 0.5_0.2
5	250-1250 µL	Finntip	12500 µL	250_1250	± 0.6_0.6	≤ 0.4_0.2
6	500-2500 µL	Finntip	25000 µL	500_2500	± 0.5_0.5	≤ 0.5_0.3
7	1000-5000 µL	Finntip	50000 µL	1000_5000	± 0.5_0.5	≤ 0.5_0.3

Thermo - Finn timer Multi

1	50-250 µL	Finntip	8	250	± 2	≤ 2
---	-----------	---------	---	-----	-----	-----

Thermo - Finn timer Biocontrol Single Channel

1	5-40 µL	Yellow	1	5_40	± 2.5_0.6	≤ 2_0.3
2	40-200 µL	Yellow	1	40_200	± 1.1_0.6	≤ 0.8_0.3
3	200-1000 µL	Blue	1	200-1000	± 0.9_0.6	≤ 0.6_0.2
4	1000-5000 µL	Green	1	1000-5000	± 1_0.5	≤ 0.4_0.2
5	2000-10000 µL	Red	1	2000-10000	± 1_0.5	≤ 0.3_0.2

Thermo - Finn timer Biocontrol Multi Channel

1	0.5-10 µL	8	1_10	± 12_2.4	≤ 8_1.6
2	5-50 µL	8_12_16	5_50	± 5_1.5	≤ 2_0.7
3	50-300 µL	8_12	50_300	± 4.6_1	≤ 1.5_0.3
4	150-1500 µL	8	300_1500	± 4_1	≤ 2.5_0.3

