



Volume Range	Colour Code	Channels	Test Volume (μL)	Inaccuracy (%)	Imprecision (%)
--------------	-------------	----------	------------------	----------------	-----------------

#### VistaLab - Ovation Fix Volume Pipette

1	1 μL	Red	1	± 2.7	≤ 1.8
2	2 μL	Red	2	± 1.5	≤ 0.9
3	5 μL	Orange	5	± 1.5	≤ 0.6
4	10 μL	Orange	10	± 1.5	≤ 0.5
5	20 μL	Yellow	20	± 1	≤ 0.3
6	25 μL	Green	25	± 1.4	≤ 0.6
7	50 μL	Green	50	± 0.8	≤ 0.24
8	100 μL	Green	100	± 0.8	≤ 0.15
9	200 μL	Blue	200	± 0.8	≤ 0.15
10	250 μL	Purple	250	± 1.2	≤ 0.3
11	300 μL	Purple	300	± 1	≤ 0.25
12	500 μL	Purple	500	± 0.8	≤ 0.2
13	1000 μL	Purple	1000	± 0.8	≤ 0.15

#### VistaLab - Ovation Adjustable Volume Single Channel Pipette

1	0.2-2 μL	Red	1	± 9.2_7_1.5	≤ 7.6_1.8_0.9
2	1-10 μL	Orange	1	± 2.5_1.5_1	≤ 1.2_0.6_0.4
3	2-20 μL	Yellow	1	± 7_1.5_1	≤ 2_0.5_0.3
4	10-100 μL	Green	1	± 3.5_0.8_0.8	≤ 1_0.24_0.15
5	20-200 μL	Blue	1	± 2.5_0.8_0.8	≤ 1_0.25_0.15
6	100-1000 μL	Purple	1	± 3_0.8_0.8	≤ 0.6_0.2_0.15

#### VistaLab - Ovation Adjustable Volume Multi Channel Pipette

1	0.5-20 μL	Yellow	8_12	± 3_1.5_1	≤ 1.8_0.8_0.4
2	2-125 μL	Green	8_12	± 3_1_0.8	≤ 1.25_0.3_0.25
3	5-250 μL	Blue	8_12	± 2.5_1_0.8	≤ 1_0.4_0.25
4	25-850 μL	Gray	12	± 3.5_1.25_1	≤ 1_0.4_0.3
5	25-1250 μL	Purple	8	± 3.5_1_0.8	≤ 0.8_0.3_0.2

#### VistaLab - Ovation Electronic Single Channel Pipette

1	0.5-20 μL	Yellow	1	± 3_1.5_1	≤ 1.8_0.5_0.3
2	2-125 μL	Green	1	± 2.9_0.8_0.8	≤ 1_0.24_0.15
3	5-250 μL	Blue	1	± 2.5_0.8_0.8	≤ 1_0.25_0.15
4	25-1250 μL	Purple	1	± 2.8_0.8_0.75	≤ 0.6_0.2_0.15
5	100-5000 μL	Gray	1	± 2.4_0.6_0.6	≤ 0.6_0.2_0.16
6	0.2-10 mL	Red	1	± 2.4_1_0.6	≤ 0.3_0.2_0.16

