

TABLE 3
VALUES FOR THE CONVERSION FACTOR Z ($\mu\text{L}/\text{mg}$)
FOR DISTILLED WATER AS A FUNCTION OF
TEMPERATURE AND ALTITUDE

Temperature °C	Altitude m			
	0	450	900	1 400
15	1.002 0	1.001 9	1.001 9	1.001 8
15.5	1.002 0	1.002 0	1.001 9	1.001 9
16	1.002 1	1.002 1	1.002 0	1.002 0
16.5	1.002 2	1.002 2	1.002 1	1.002 0
17	1.002 3	1.002 2	1.002 2	1.002 1
17.5	1.002 4	1.002 3	1.002 3	1.002 2
18	1.002 5	1.002 4	1.002 4	1.002 3
18.5	1.002 6	1.002 5	1.002 5	1.002 4
19	1.002 7	1.002 6	1.002 5	1.002 5
19.5	1.002 8	1.002 7	1.002 6	1.002 6
20	1.002 9	1.002 8	1.002 7	1.002 7
20.5	1.003 0	1.002 9	1.002 8	1.002 8
21	1.003 1	1.003 0	1.003 0	1.002 9
21.5	1.003 2	1.003 1	1.003 1	1.003 0
22	1.003 3	1.003 2	1.003 2	1.003 1
22.5	1.003 4	1.003 3	1.003 3	1.003 2
23	1.003 5	1.003 5	1.003 4	1.003 3
23.5	1.003 6	1.003 6	1.003 5	1.003 5
24	1.003 8	1.003 7	1.003 6	1.003 6
24.5	1.003 9	1.003 8	1.003 8	1.003 7
25	1.004 0	1.003 9	1.003 9	1.003 8
25.5	1.004 1	1.004 1	1.004 0	1.004 0
26	1.004 3	1.004 2	1.004 2	1.004 1
26.5	1.004 4	1.004 3	1.004 3	1.004 2
27	1.004 5	1.004 5	1.004 4	1.004 4
27.5	1.004 7	1.004 6	1.004 6	1.004 5
28	1.004 8	1.004 8	1.004 7	1.004 6
28.5	1.005 0	1.004 9	1.004 8	1.004 8
29	1.005 1	1.005 0	1.005 0	1.004 9
29.5	1.005 2	1.005 2	1.005 1	1.005 1
30	1.005 4	1.005 3	1.005 3	1.005 2